radwag.com

100 µl ÷ 1000 µl - RW8-106-20-9 Automatic Pipettes

PA-203-0006

More information on the website
mirror.radwag.com/en/info,w1,9XE



The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

Metrological parameters	
Volume	100 – 1000 µl
Systematic error	2.0% / 0.6%
Random error	0.7% / 0.2%
Tips	1000 µl

Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration
MYA 5Y.P Microbalance for pipettes calibration
Automatic Balances for Pipettes Calibration