



**EU DECLARATION OF CONFORMITY (No. 35/1/18/EN)**

1	Non-automatic weighing instrument Instrument model (the medical device): Baby scales <b>WPT 6/15D ; WPT 10/20D</b> Personal scales <b>WPT 60/150 O ; WPT 60/150 OW ;</b> <b>WPT 100/200 O ; WPT 100/200 OW</b> Bed scales <b>WPT/4B 500C ; WPT/8B 300C</b> Chair scales <b>WPT/4K 150C ; WPT/K 250C</b>	
2	Manufacturer:	RADWAG WAGI ELEKTRONICZNE Witold Lewandowski
	Address:	POLAND, 26-600 Radom, 5 Toruńska Street
3	This declaration of conformity is issued under the sole responsibility of the manufacturer	
4	Class I with measuring function, rule 12 Annex IX of the 93/42/EEC directive. Conformity assessment procedure according to MDD directive – Annex VII + Annex V - measuring function	
5	The object of the declaration described above is in conformity with the relevant Union harmonisation legislation:	
6	Directive	Harmonised Standards
	2014/31/EU	EN 45501
	93/42/EEC	
7	Notified Body no. 1383, Czech Metrology Institute, performed EC type-examination and issued the certificate: EC type-approval certificate no. TCM 128/14-5190. Notified Body no. 1383, Czech Metrology Institute, performed examination and issued the Test certificate no. ZR 128/06-0050 for measuring indicator.	
8	Additional information: The manufacturer's Quality System accordant with module D is under the supervision of Notified Body no. 1383, Czech Metrology Institute (Directive NAWI). The manufacturer's Quality System is under the supervision of Notified Body no. 2274, TÜV NORD Polska Sp. z o.o. (Directive MDD)	

Signed for and on behalf of:

The Managing Director of RADWAG WAGI ELEKTRONICZNE – *Dr. Witold Lewandowski*

Radom 08.02.2018

.....  
*Place and date of issue*

Jacek Pilecki, *Manager of Quality Control Department*

.....  
*Name, Position*

  
.....  
*Signature*